



## **CERTIFICATE OF ANALYSIS**

Product: **EDTA, 0.050 Molar**  
Catalog Number: E1006  
Lot Number: E10061712501  
Manufacture Date: 06/20/25  
Expiration Date: 06/20/27

<b>Property</b>	<b>Specification</b>	<b>Result</b>
Assay, EDTA	0.0500 ± 0.0008 M	0.0500 M

Prepared with ACS Reagent Grade ethylenediaminetetraacetic acid, disodium salt, dihydrate, and high purity water. Standardized against calcium carbonate standard traceable to NIST. Indicator: Eriochrome Black T. As part of final quality control release, each lot is assayed in duplicate to confirm the concentration value.

Certificate approved on 06/23/25 by:



Grant Christen  
Quality Systems and Regulatory Affairs  
Rocky Mountain Reagents, Inc.