



CERTIFICATE OF ANALYSIS

Product: **Zinc Standard, 1 mg/L in 1% w/w Hydrochloric Acid**
Catalog Number: Z1015
Lot Number: Z10152932501
Manufacture Date: 10/20/25
Expiration Date: 10/20/26

Property	Specification	Result
Concentration, Zinc	1.00 ± 0.02 mg/L	Conforms

Prepared from 1000 ppm AA zinc standard (certified traceable to NIST), ACS Reagent Grade hydrochloric acid, and deionized water. Dilutions were made with Class A volumetric glassware.

Certificate approved on 10/20/25 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.