




CERTIFICATE OF ANALYSIS

Product: **Hydrochloric Acid, 0.100 Normal in Isopropyl Alcohol**
Catalog Number: H1040
Lot Number: H10402972501
Manufacture Date: 10/24/25
Expiration Date: 10/24/27

Property	Specification	Result
Assay, HCl	0.1000 ± 0.0008 N	0.0999 N

Prepared with ACS Reagent Grade hydrochloric acid and ACS Reagent Grade isopropyl alcohol. Standardized against sodium hydroxide solution certified traceable to NIST. Indicator: phenolphthalein. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 10/24/25 by:



Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.