



CERTIFICATE OF ANALYSIS

Product: **Potassium Permanganate, 0.200 Normal**
Catalog Number: P1072
Lot Number: P10723282501
Manufacture Date: 11/24/25
Expiration Date: 11/24/27

Property	Specification	Result
Assay, KMnO ₄	0.200 ± 0.005 N	0.200 N

Prepared with ACS Reagent Grade potassium permanganate and high purity water. Standardized against sodium oxalate solution certified traceable to NIST. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 11/24/25 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.