



CERTIFICATE OF ANALYSIS

Product: **Potassium Hydroxide, 1.00 Normal**
Catalog Number: P1042
Lot Number: P10420122601
Manufacture Date: 01/12/26
Expiration Date: 01/12/28

Property	Specification	Result
Assay, KOH	1.000 ± 0.008 N	1.000 N

Prepared with reagent grade potassium hydroxide and high purity water. Standardized against hydrochloric acid solution certified traceable to NIST. Indicator: phenolphthalein. As part of final QC release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 01/12/26 by:

Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.