



CERTIFICATE OF ANALYSIS

Product: **EDTA, 0.0575 Molar**
Catalog Number: E1007
Lot Number: E10070972601
Manufacture Date: 04/07/26
Expiration Date: 04/07/28

Property	Specification	Result
Assay, EDTA	0.0575 ± 0.0008 M	0.0576 M

Prepared with ACS Reagent Grade ethylenediaminetetraacetic acid, disodium salt, dihydrate, and high purity water. Standardized against calcium carbonate standard certified traceable to NIST. Indicator: Eriochrome Black T. As part of final quality control release, this lot was assayed in duplicate to confirm the concentration value.

Certificate approved on 04/13/26 by:

Matt Jones
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.