Rocky Mountain Reagents
SCIENTIFIC SOLUTIONS SINCE 1951

CERTIFICATE OF ANALYSIS

Product:	**VIP Stain (Vaginal Identification of Pathogens)**
Catalog Number:	V1000
Lot Number:	V10001602601
Manufacture Date:	06/09/26
Expiration Date:	06/09/27

Property	Specification	Result
Concentration, Crystal Violet	0.40 ± 0.01 g/L	Conforms
pH, Phosphate Buffer Solvent at 25 °C	6.0 ± 0.1	Conforms

Prepared with BSC certified crystal violet, ACS Reagent Grade sodium phosphate, dibasic, ACS Reagent Grade potassium phosphate, monobasic, and high purity water.

Certificate approved on 06/10/26 by:



Grant Christen
Quality Systems and Regulatory Affairs
Rocky Mountain Reagents, Inc.